

November 5, 2014

URGENT RECALL

Carefusion Chlorhexidine Gluconate (CHG) Surgical Scrub Brush

Dear Valued Life-Assist Customer:

Carefusion is recalling ALL LOTS of 4458A CHG Surgical Scrub Brush due to higher concentrations of available CHG than what is stated on the product labeling.

Our records indicate you have purchased this product from Life-Assist, Inc.

Life-Assist Item Numbers	
MEDL_CFN4458AZ	SCRUB BRUSH, 4% CHG, BX/30
PHS_371073	SCRUB BRUSH, EZ SCRUB 4% CHG, BX/30

Please complete the acknowledgement/certificate of destruction form and email to customerservice@life-assist.com, or fax to Life-Assist Customer Service at 800-290-9794. The form must be completed and returned, even if you have no stock on hand.

Credit will be issued to Life-Assist for the amount indicated on the form, which we will pass on to you the customer.

Please contact Life-Assist, Inc. Customer Service at 800-824-6016 or customerservice@life-assist.com with any questions.

We apologize for any inconvenience.



**Acknowledgement/Certificate of destruction for
Carefusion Chlorhexidine Gluconate (CHG) Surgical Scrub Brushes**

Please complete this form and email to customerservice@life-assist.com, or fax to Life-Assist Customer Service at 800-290-9794. The form must be completed and returned, even if you have no stock on hand.

Manufacturer Item Number	Life-Assist Item Number (used to purchase product)	Quantity Destroyed
4458A	MEDL_CFN4458AZ	
4458A	PHS_371073	

***Please follow your facility's destruction protocol/procedure when destroying this or any product.**

_____ We do not have any of these recalled items in our possession.

I certify that the product identified above has been destroyed.

_____ Account#

_____ Facility Name

_____ Print Name

_____ Title

_____ Signature _____ Date

_____ Email Address

_____ Telephone Number

_____ Fax Number

